

---

## **Karl Fischer Titration Tester (all automatic operation)**

**BGD 232**

---

### **Product Description**

These instruments use Kari-Fischer measurement to measure the trace of moisture in the coating, ink etc. It use ascertaining end-point titrimetry to measure the end-point. It can be used to measure the moisture content in chemical fertilizer medicine, foodstuff, chemical material and other industrial products.

It can absorb and transfuse liquid automatically measure the end-point and drain waste liquid automatically, and show titrimetry value by LED digital display. It has many advatages such as easy to operate, convenient to use and reliable test results etc.



---

### **Technical Specification**

- Polarization Voltage: -20mv
- Sensitivity: 10-6A
- Distinguish Ability: 0.01ml
- End-point Delay: 10±2 second
- Titration Flux: control automatically
- Power Supply: 220V /50Hz
- Power: 18W

# **Karl Fischer Titration Tester (all automatic operation)**

**BGD 232**

## **Main Technical Parameters**

---

Ordering information BGD 232 - Karl Fischer Titration Tester

## **Disclaimer**

---

The information given in this sheet is not intended to be exhaustive and any person using the product for any purpose other than that specifically recommended in this sheet without first obtaining written confirmation from us as to the suitability of the product for the intended purpose does so at his own risk. Whilst we endeavour to ensure that all advice we give about the product (whether in this sheet or otherwise) is correct we have no control over either the quality or condition of the product or the many factors affecting the use and application of the product. Therefore, unless we specifically agree in writing to do so, we do not accept any liability whatsoever or howsoever arising for the performance of the product or for any loss or damage (other than death or personal injury resulting from our negligence) arising out of the use of the product. The information contained in this sheet is liable to modification from time to time in the light of experience and our policy of continuous product development