

Moisture analyzer (Max 110 g: d=0,001 g)


DAB100-3

Product Description

Particularly user-friendly moisture analyzer with high-quality halogen quartz glass heater

Technical Data

| Parameter | Value |
|------------------------------------|--|
| Weighing capacity [Max] | 110 g |
| Readability [d] | 0,001 g |
| Weighing system | Strain gauge |
| Temperature [Min] (°C) | 40 °C |
| Temperature [Max] (°C) | 199 °C |
| Stabilization time | 3 s |
| Units | g, % Humidity, % Dry mass |
| Reproducibility | 0,001 g |
| Linearity | ± 0,003 g |
| Readability moisture [d] (%) | 0,01% |
| Drying modes | Standard drying, Rapid drying, Gentle drying |
| Switch-off criteria | time controlled, automatic, manual |
| Input voltage device / power [Max] | 230 V AC, 50 Hz |
| Product group | Moisture analyzer |



Moisture analyzer (Max 110 g: d=0,001 g)

DAB100-3

Moisture analyzer (Max 110 g: d=0,001 g)

DAB100-3

Main Technical Parameters

Ordering information: DAB 100-3 Moisture Analyzer

Disclaimer

The information given in this sheet is not intended to be exhaustive and any person using the product for any purpose other than that specifically recommended in this sheet without first obtaining written confirmation from us as to the suitability of the product for the intended purpose does so at his own risk. Whilst we endeavour to ensure that all advice we give about the product (whether in this sheet or otherwise) is correct we have no control over either the quality or condition of the product or the many factors affecting the use and application of the product. Therefore, unless we specifically agree in writing to do so, we do not accept any liability whatsoever or howsoever arising for the performance of the product or for any loss or damage (other than death or personal injury resulting from our negligence) arising out of the use of the product. The information contained in this sheet is liable to modification from time to time in the light of experience and our policy of continuous product development